



## **A wearable microwave detector for diagnosing thoracic injuries-test on a porcine pneumothorax model**

Downloaded from: <https://research.chalmers.se>, 2023-05-05 12:34 UTC

Citation for the original published paper (version of record):

Oveland, N., Buendia, R., Sjöqvist, B. et al (2015). A wearable microwave detector for diagnosing thoracic injuries-test on a porcine pneumothorax model. *Scandinavian Journal of Trauma, Resuscitation and Emergency Medicine*, 23(2).  
<http://dx.doi.org/10.1186/1757-7241-23-S2-A20>

N.B. When citing this work, cite the original published paper.

MEETING ABSTRACT

Open Access

# A wearable microwave detector for diagnosing thoracic injuries-test on a porcine pneumothorax model

Nils Petter Oveland<sup>1,2,3\*</sup>, Ruben Buendia<sup>4,5,6</sup>, Bengt Arne Sjöqvist<sup>4,5,6</sup>, Marianne Oropeza-Moe<sup>7</sup>, Nina Gjerde Andersen<sup>2</sup>, Andreas Fager<sup>4,5</sup>, Mikael Persson<sup>4,5</sup>, Mikael Elam<sup>5,8</sup>, Stefan Candefjord<sup>4,5,6</sup>

From London Trauma Conference 2014  
London, UK. 9-12 December 2014

## Background

In the prehospital setting, a point-of-care diagnostic test is needed to diagnose pneumothorax (PTX) and monitor its progression to prevent unnecessary patient morbidity and mortality. Ultrasonography is more sensitive than supine chest x-ray for diagnosing PTX, but the accuracy depends on the experience of the operator. Therefore, a non-operator dependent instrument would be valuable for detection and continuous monitoring of an evolving PTX.

## Study objective

To evaluate the potential of a new microwave technology for diagnosing PTX.

## Methods

An experimental PTX model was set up in two anesthetized pigs. A belt with eight microwave antennas was strapped around the pig's chest. Air was insufflated into a catheter in the right pleural space in twelve incremental steps (PTX volumes: 50, 100, 150, 200, 250, 300, 400, 500, 750, 1000, 1500, 2000 mL). Each injection was followed by a measurement with the microwave detector. A computer-based classification algorithm was used to distinguish between the measurements using a leave-one-out approach (i.e. the sample to be classified was not included in the training data matrix), where each PTX volume was treated as an individual class.

## Results

The microwave belt was able to differentiate between normal lungs and PTX in both animals with an overall diagnostic accuracy of 100 % (i.e. a sensitivity and specificity of 100%). Furthermore, the classification accuracy for predicting the size of PTX was 100% and 98% for each pig, respectively.

## Conclusion

The microwave technology proved promising in diagnosing and predicting size of PTX ranging from 50 mL to 2000 mL. This within-model experiment only differentiated PTX and normal lungs in individual pigs and not between different animals. A larger validation study needs to be done to further evaluate the diagnostic accuracy of the microwave detector.

## Institution

This preliminary study was conducted at Sandnes Education and Research Center Høyland (SEARCH), Sandnes, Norway.

## Conflict of interest

The authors state no conflicts of interest.

## Authors' details

<sup>1</sup>Department of Health Studies, Network for Medical Sciences, University of Stavanger, Stavanger, Norway. <sup>2</sup>Department of Anesthesiology and Intensive Care, Stavanger University Hospital, Stavanger, Norway. <sup>3</sup>Department of Research and Development, Norwegian Air Ambulance Foundation, Drøbak, Norway. <sup>4</sup>Chalmers University of Technology, Gothenburg, Sweden. <sup>5</sup>MedTech West, Gothenburg, Sweden. <sup>6</sup>SAFER Vehicle and Traffic Safety Centre at Chalmers University of Technology, Gothenburg, Sweden. <sup>7</sup>Department of Production Animal Clinical Sciences, Faculty of Vetbio, Norwegian University of Life Sciences, Sandnes, Norway. <sup>8</sup>Sahlgrenska University Hospital, Gothenburg, Sweden.

\* Correspondence: nils.petter.oveland@me.com

<sup>1</sup>Department of Health Studies, Network for Medical Sciences, University of Stavanger, Stavanger, Norway.

Full list of author information is available at the end of the article

Published: 11 September 2015

doi:10.1186/1757-7241-23-S2-A20

**Cite this article as:** Oveland et al.: A wearable microwave detector for diagnosing thoracic injuries-test on a porcine pneumothorax model. *Scandinavian Journal of Trauma, Resuscitation and Emergency Medicine* 2015 **23**(Suppl 2):A20.

**Submit your next manuscript to BioMed Central and take full advantage of:**

- Convenient online submission
- Thorough peer review
- No space constraints or color figure charges
- Immediate publication on acceptance
- Inclusion in PubMed, CAS, Scopus and Google Scholar
- Research which is freely available for redistribution

Submit your manuscript at  
[www.biomedcentral.com/submit](http://www.biomedcentral.com/submit)

